



MICROGEN BIOPRODUCTS
CERTIFICATE OF CONFORMITY

Name of Product: Microgen Streptococcus ID 20 test
Product Code: MID-62
Components: 20 x MID 62C; 20 x MID62B 1 x MID62D

1. All components in this kit comply with current Microgen Bioproducts approved specifications.

2. Performance of strips with QC test panel complies with current release specifications:

Streptococcus mutans MBCC 1067

Streptococcus uberis MBCC 1072

Gemella haemolysans MBCC 1085

Enterococcus avium MBCC 1136

Enterococcus faecalis MBCC 1137

Give expected pattern of substrate reactions.

3. Product satisfies final Quality Control examination according to FPS MID-62

Component	Inventory Code	Lot Number	Expiry	Lot-expiry correct SOP340	Component PMP reviewed SOP337
Suspending Broth	MID62B	2501	2018-07	PASS	PASS
Test Strip	MID62C	2501	2018-04	PASS	PASS
Hippurate	MID62D	2601	2018-04	PASS	PASS
Assembly	MID62	02561	2018-04	PASS	PASS
Overall Expiry Date:			2018-04		

SIGNATURE: RAJ

DATE: 15 / NOV / 2016